


Programmer's Guide

JETI Software Development Kit jeti_radio_ex.dll

Version 4.8.x



**JETI Technische Instrumente GmbH
Göschwitzer Str. 48
D-07745 Jena**

Tel.: +49-3641-2329200

Fax: +49-3641-2329201

e-mail: sales@jeti.com

internet: www.jeti.com

Table of contents

1	JETI SDK Overview	3
2	Introduction	4
2.2	How to communicate	4
3	Function Reference	5
3.1	JETI_GetRadioExDLLVersion	6
3.2	JETI_GetNumRadioEx	7
3.3	JETI_GetSerialRadioEx	8
3.4	JETI_OpenRadioEx	9
3.5	JETI_CloseRadioEx	10
3.6	JETI_MeasureEx	11
3.7	JETI_PrepareMeasureEx	12
3.8	JETI_MeasureStatusEx	13
3.9	JETI_MeasureBreakEx	14
3.10	JETI_SpecRadEx	15
3.11	JETI_SaveSpecRadSPCEx	16
3.12	JETI_SaveSpecRadCSVEx	17
3.13	JETI_SpecRadHiResEx	18
3.14	JETI_RadioEx	19
3.15	JETI_PhotoEx	20
3.16	JETI_ChromxyEx	21
3.17	JETI_Chromxy10Ex	22
3.18	JETI_ChromuvEx	23
3.19	JETI_ChromXYZEx	24
3.20	JETI_DWLPEEx	25
3.21	JETI_CCTEx	26
3.22	JETI_DuvEx	27
3.23	JETI_CRIEx	28
3.24	JETI_PeakFWHMEx	29
3.25	JETI_BlueMeasurementEx	30
3.26	JETI_RadioTintEx	31
3.27	JETI_SetMeasDistEx	32
3.28	JETI_GetMeasDistEx	33
4	Examples	34
4.2	C Examples	34
4.3	LabVIEW Examples	34
4.4	VisualBasic / VBA Examples	34
5	Appendix A	35
6	License Agreement	36
7	Service	37

1 JETI SDK Overview

The JETI Software Development Kit provides a complete software solution for interfacing spectrometric and radiometric devices from JETI Technische Instrumente GmbH. No firmware command expertise is required. Instead, a simple, high-level Application Program Interface (API) is used to provide complete connectivity. The API is provided in the form of several Windows Dynamic Link Libraries (DLL). The libraries can be used by any programming language that can handle DLL's such C/C++, VisualBasic, or LabVIEW. To get access to the functions the needed DLL files must be copied to the Windows System Folder or to the working directory of the calling application.

The following DLLs are available:

- 1 jeti_spectro.dll / jeti_spectro64.dll
1.2 provides a set of functions for simple spectrometric measurement
- 2 jeti_spectro_ex.dll / jeti_spectro_ex64.dll
2.2 a set of functions like jeti_spectro.dll, but with more options to control the measurement
- 3 jeti_radio.dll / jeti_radio64.dll
3.2 provides a set of functions for simple radiometric measurement, including calculation of colorimetric values (e.g. xy- and u'v'-values, CCT, CRI,...)
- 4 jeti_radio_ex.dll / jeti_radio_ex64.dll
4.2 a set of functions like jeti_radio.dll, but with more options to control the measurement and calculations
- 5 jeti_core.dll / jeti_core64.dll
5.2 a set of functions to fully control the device and perform custom measurement sequences

Please note that this documentation describes only the functions provided by the jeti_radio_ex.dll. For description of the other DLL's please refer to the corresponding documents.

2 Introduction

The jeti_radio_ex API is provided in the form of a Windows Dynamic Link Library (DLL). The interface DLL communicates with the device via the provided device driver and the basic driver DLL jeti_core.dll.

JETI Technische Instrumente GmbH offers two versions of the DLL. The first version is for 32bit Windows operating systems (Windows 10/11).

The second version is for real 64-bit programs under the 64 bit versions of Windows 10/11.

There are no differences in the functionality between the two versions.

2.2 How to communicate

To get access to the functions you must copy the files jeti_radio_ex.dll and jeti_core.dll to the working directory of your application, or to the windows\system32 directory.

In general, the user initiates communication with the target device(s) by making a call to JETI_GetNumRadioEx. This call will return the number of target devices. This number is then used as a range when calling JETI_GetSerialRadioEx to build a list of device serial numbers.

To access a device, it must first be opened by a call to JETI_OpenRadioEx using an index determined from the call to JETI_GetNumRadioEx. The JETI_OpenRadioEx function will return a handle to the device that is used in all subsequent accesses. When I/O operations are complete, the device is closed by a call to JETI_CloseRadioEx.

In case of a fatal communication error (error code 0xFF) JETI_HardReset (from jeti_core.dll) could be used to reset the device and resume the communication. For more information see the function description of JETI_HardReset in 'JETI SDK Programmer's Guide jeti_core.dll' and the Appendix A.

3 Function Reference

Convention for calling: `__stdcall`

Type	Size in Bit	Minimum	Maximum
DWORD (unsigned long)	32	0	$2^{32}-1$
DWORD_PTR (unsigned long integer)	32 (32bit DLLs)	0	$2^{32}-1$
(unsigned long long)	64 (64bit DLLs)	0	$2^{64}-1$
WORD (unsigned short integer)	16	0	65535
FLOAT (IEEE standard)	32	-3.40282E+38	3.40282E+38
BOOL (integer)	32	-2^{21}	$2^{31}-1$

3.1 JETI_GetRadioExDLLVersion

This function returns the current version number of the jeti_radio_ex DLL.

Prototype

DWORD JETI_GetRadioExDLLVersion (WORD *wMajorVersion, WORD *wMinorVersion, WORD *wBuildNumber)

Parameters

Input

Name	Type	Description	Call
wMajorVersion	WORD *	address of a WORD variable that will contain the major version	By reference
wMinorVersion	WORD *	address of a WORD variable that will contain the minor version	By reference
wBuildNumber	WORD *	address of a WORD variable that will contain the build number	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.2 JETI_GetNumRadioEx

This function returns the number of JETI devices connected to the PC or to the LAN.

Prototype

DWORD JETI_GetNumRadioEx (DWORD *dwNumDevices)

Parameters

Input

Name	Type	Description	Call
dwNumDevices	DWORD *	address of a DWORD variable that will contain the number of devices connected	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.3 JETI_GetSerialRadioEx

This function returns the serial numbers for the device specified by an index passed in dwDeviceNum. The index for the first device is 0 and the last device is the value returned by JETI_GetNumRadioEx – 1.

Prototype

DWORD JETI_GetSerialRadioEx (DWORD dwDeviceNum, char *cBoardSerialNr, char *cSpecSerialNr, char *cDeviceSerialNr)

Parameters

Input

Name	Type	Description	Call
dwDeviceNum	DWORD	index of the device for which the serial numbers are desired	By value
cBoardSerialNr	char *	address of a string that will contain the electronics serial number	By reference
cSpecSerialNr	char *	address of a string that will contain the spectrometer serial number	By reference
cDeviceSerialNr	char *	address of a string that will contain the device serial number	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.4 JETI_OpenRadioEx

Opens a device (using device number returned by JETI_GetNumRadioEx) and returns a handle which will be used for subsequent accesses.

Prototype

DWORD JETI_OpenRadioEx (DWORD dwDeviceNum, DWORD_PTR *dwDevice)

Parameters

Input

Name	Type	Description	Call
dwDeviceNum	DWORD	Device index. 0 for first device, 1 for second, etc.	By value
dwDevice	DWORD_PTR *	Pointer to a variable where the handle to the device will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.5 JETI_CloseRadioEx

Closes an open device using the handle provided by JETI_OpenRadioEx.

Prototype

DWORD JETI_CloseRadioEx (DWORD_PTR dwDevice)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device to close as returned by JETI_OpenRadioEx	By value

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS
	0x... see Appendix A for error codes

3.6 JETI_MeasureEx

Starts a radiometric measurement.

The integration time will be determined automatically if fTint is set to zero (0). Otherwise the integration time in milliseconds will be used.

To perform an average measurement with up to 10000 measurements set the wAver to the appropriate value, otherwise set the value to 1.

All further calculations will be performed using the wavelength resolution as specified by the dwStep value. A value of 1 or 5 nm is allowed.

NOTE: The function will return *immediately*. Before any other DLL-function call the function JETI_MeasureStatusEx must be used to check if the measurement has finished.
Please note that a measurement could take several seconds up to 2 minutes, depending on the intensity of the light source to measure. If averaging is used, the time increases proportional with the count.

Prototype

DWORD JETI_MeasureEx (DWORD_PTR dwDevice, float fTint, WORD wAver, DWORD dwStep)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fTint	float	Integration time in [ms] for the measurement (up to 60000). Set this value to 0 for adaption.	By value
wAver	WORD	Count of scans for averaging. Set this value to 1 for no averaging	By value
dwStep	DWORD	wavelength step in [nm] for further calculations	By value

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.7 JETI_PrepareMeasureEx

Prepares a radiometric measurement.

The integration time will be determined automatically if fTint is set to zero (0). Otherwise the integration time in milliseconds will be used.

To perform an average measurement with up to 10000 measurements set the wAver to the appropriate value, otherwise set the value to 1.

All further calculations will be performed using the wavelength resolution as specified by the dwStep value. A value of 1 or 5 nm is allowed.

It will not start the measurement.

NOTE: The function will return **immediately**. Before any other DLL-function call the function JETI_MeasureStatusEx must be used to check if the measurement has started by an external trigger.
Please note that a measurement could take several seconds up to 2 minutes, depending on the intensity of the light source to measure. If averaging is used, the time increases proportional with the count.

Prototype

DWORD JETI_PrepareMeasureEx (DWORD_PTR dwDevice, float fTint, WORD wAver, DWORD dwStep)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fTint	float	Integration time in [ms] for the measurement (up to 60000). Set this value to 0 for adaption.	By value
wAver	WORD	Count of scans for averaging. Set this value to 1 for no averaging	By value
dwStep	DWORD	wavelength step in [nm] for further calculations	By value

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.8 JETI_MeasureStatusEx

Returns the status of a measurement started with JETI_MeasureEx. A measurement has finished if the boStatus variable is FALSE (0). If the measurement is already in progress the variable boStatus returns TRUE (1).

If a measurement was initiated with automatic adaption of integration time and the measurement could not be performed because of overexposure boStatus will be switched to FALSE (0) and the function will return an error code 0x20.

NOTE: A function to get a measuring result should not be called until the JETI_MeasureStatusEx returns that the measurement has finished.

Prototype

DWORD JETI_MeasureStatusEx (DWORD_PTR dwDevice, BOOL *boStatus)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
boStatus	BOOL *	Pointer to a variable where the status will be stored TRUE (1) – in progress FALSE (0) – ready	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.9 JETI_MeasureBreakEx

This function cancels an initiated measurement.

Prototype

DWORD JETI_MeasureBreakEx (DWORD_PTR dwDevice)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.10 JETI_SpecRadEx

Returns the spectral radiance per wavelength. The wavelength range in [nm] is specified by dwBeg and dwEnd. The wavelength step was specified by the function JETI_MeasureEx.

Please note that the array for the values must provide enough space for at least the count of calculated values! The count of values can be obtained by subtract the start wavelength (dwBeg) from the end wavelength (dwEnd), divide the result by the wavelength step and increase the result by 1:

$$count = \frac{dwEnd - dwBeg}{dwStep} + 1$$

The unit of the values depends on the used measuring head. By default it's as follows:

Measuring Head	Description	Unit
none	spectral radiance	$\frac{W}{sr \times m^2 \times nm}$
cosine corrector head-piece	spectral irradiance	$\frac{W}{m^2 \times nm}$
integrating sphere	spectral radiant flux	$\frac{W}{nm}$
tube	spectral radiant intensity	$\frac{W}{sr \times nm}$

Prototype

DWORD JETI_SpecRadEx (DWORD_PTR dwDevice, DWORD dwBeg, DWORD dwEnd, FLOAT *fSprad)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
dwBeg	DWORD	start wavelength in [nm]	By value
dwEnd	DWORD	end wavelength in [nm]	By value
fSprad	FLOAT *	Pointer to an array where the spectral radiance values will be stored (the array must provide space for at least 81 values)	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x.. see Appendix A for error codes

3.11 JETI_SaveSpecRadSPCEx

Writes the spectral radiance per wavelength to a SPC-file. The wavelength range in [nm] is specified by dwBeg and dwEnd. The wavelength step was specified by the function JETI_MeasureEx.

The unit of the values depends on the used measuring head. By default it's as follows:

Measuring Head	Description	Unit
none	spectral radiance	$W / sr \times m^2 \times nm$
cosine corrector head-piece	spectral irradiance	$W / m^2 \times nm$
integrating sphere	spectral radiant flux	W / nm
tube	spectral radiant intensity	$W / sr \times nm$

Prototype

DWORD JETI_SaveSpecRadSPCEx (DWORD_PTR dwDevice, DWORD dwBeg, DWORD dwEnd, char *cPathName, char *cOperator, char *cMemo)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
dwBeg	DWORD	start wavelength in [nm]	By value
dwEnd	DWORD	end wavelength in [nm]	By value
cPathName	char *	String (up to 255 character) with file name and directory to save the file (file extension should be .SPC)	By reference
cOperator	char *	String (up to 63 character) with operator's name	By reference
cMemo	char *	String (up to 129 character) with a measurement description	By Reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x.. see Appendix A for error codes

3.12 JETI_SaveSpecRadCSVEx

Writes the spectral radiance per wavelength to a CSV-file. The wavelength range in [nm] is specified by dwBeg and dwEnd. The wavelength step was specified by the function JETI_MeasureEx.

The unit of the values depends on the used measuring head. By default it's as follows:

Measuring Head	Description	Unit
none	spectral radiance	$W / sr \times m^2 \times nm$
cosine corrector head-piece	spectral irradiance	$W / m^2 \times nm$
integrating sphere	spectral radiant flux	W / nm
tube	spectral radiant intensity	$W / sr \times nm$

Prototype

DWORD JETI_SaveSpecRadCSVEx (DWORD_PTR dwDevice, DWORD dwBeg, DWORD dwEnd, char *cPathName, char *cOperator, char *cMemo)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
dwBeg	DWORD	start wavelength in [nm]	By value
dwEnd	DWORD	end wavelength in [nm]	By value
cPathName	char *	String (up to 255 character) with file name and directory to save the file (file extension should be .CSV)	By reference
cOperator	char *	String (up to 63 character) with operator's name	By reference
cMemo	char *	String (up to 129 character) with a measurement description	By Reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x.. see Appendix A for error codes

3.13 JETI_SpecRadHiResEx

Returns the spectral radiance per wavelength. The wavelength range in [nm] is specified by dwBeg and dwEnd. The wavelength step for this function is 0.1nm.

Please note that the array for the values must provide enough space for at least the count of calculated values! The count of values can be obtained by subtract the start wavelength (dwBeg) from the end wavelength (dwEnd), divide the result by the wavelength step and increase the result by 1:

$$count = \frac{dwEnd - dwBeg}{0.1} + 1$$

The unit of the values depends on the used measuring head. By default, it's as follows:

Measuring Head	Description	Unit
none	spectral radiance	$W / sr \times m^2 \times nm$
cosine corrector head-piece	spectral irradiance	$W / m^2 \times nm$
integrating sphere	spectral radiant flux	W / nm
tube	spectral radiant intensity	$W / sr \times nm$

Prototype

DWORD JETI_SpecRadHiResEx (DWORD_PTR dwDevice, DWORD dwBeg, DWORD dwEnd, FLOAT *fSprad)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
dwBeg	DWORD	start wavelength in [nm]	By value
dwEnd	DWORD	end wavelength in [nm]	By value
fSprad	FLOAT *	Pointer to an array where the spectral radiance values will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x.. see Appendix A for error codes

3.14 JETI_RadioEx

Returns the radiometric value determined by the last measure. The wavelength range in [nm] is specified by dwBeg and dwEnd. The wavelength step was specified by the function JETI_MeasureEx.

The unit of the value depends on the used measuring head. By default it's as follows:

Measuring Head	Description	Unit
none	radiance	$W / sr \times m^2$
cosine corrector head-piece	irradiance	W / m^2
integrating sphere	radiant flux	W
tube	radiant intensity	W / sr

Prototype

DWORD JETI_Radio (DWORD_PTR dwDevice, DWORD dwBeg, DWORD dwEnd, FLOAT *fRadio)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
dwBeg	DWORD	start wavelength in [nm]	By value
dwEnd	DWORD	end wavelength in [nm]	By value
fRadio	FLOAT *	Pointer to a variable where the radiometric value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.15 JETI_PhotoEx

Returns the photometric value determined by the last measure. The wavelength range for calculation is 380...780 nm. The wavelength step was specified by the function JETI_MeasureEx.

The unit of the value depends on the used measuring head. By default it's as follows:

Measuring Head	Description	Unit
none	luminance	cd/m^2
cosine corrector head-piece	illuminance	lx
integrating sphere	luminous flux	lm
tube	luminous intensity	cd

Prototype

DWORD JETI_Photo (DWORD_PTR dwDevice, FLOAT *fPhoto)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fPhoto	FLOAT *	Pointer to a variable where the photometric value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.16 JETI_ChromxyEx

Returns the CIE-1931 chromaticity coordinates x and y determined by the last measure. The calculation is based on a 2° observer, and the wavelength range for calculation is 380...780 nm.

Prototype

DWORD JETI_ChromxyEx (DWORD_PTR dwDevice, FLOAT *fChromx, FLOAT *fChromy)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fChromx	FLOAT *	Pointer to a variable where the x-value will be stored	By reference
fChromy	FLOAT *	Pointer to a variable where the y-value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.17 JETI_Chromxy10Ex

Returns the CIE-1931 chromaticity coordinates x and y determined by the last measure. The calculation is based on a 10° observer, and the wavelength range for calculation is 380...780 nm.

Prototype

DWORD JETI_Chromxy10Ex (DWORD_PTR dwDevice, FLOAT *fChromx10, FLOAT *fChromy10)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fChromx10	FLOAT *	Pointer to a variable where the x-value will be stored	By reference
fChromy10	FLOAT *	Pointer to a variable where the y-value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.18 JETI_ChromuvEx

Returns the CIE-1976 chromaticity coordinates u' and v' determined by the last measure. The calculation is based on a 2° observer, and the wavelength range for calculation is 380...780 nm.

Prototype

DWORD JETI_ChromuvEx (DWORD_PTR dwDevice, FLOAT *fChromu, FLOAT *fChromv)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fChromu	FLOAT *	Pointer to a variable where the u' -value will be stored	By reference
fChromv	FLOAT *	Pointer to a variable where the v' -value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.19 JETI_ChromXYZEx

Returns the tristimulus XYZ determined by the last measure. The calculation is based on a 2° observer, and the wavelength range for calculation is 380...780 nm.

Prototype

DWORD JETI_ChromXYZEx (DWORD_PTR dwDevice, FLOAT *fX, FLOAT *fY, FLOAT *fZ)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fX	FLOAT *	pointer to a variable where the tristimulus X will be stored	By reference
fY	FLOAT *	pointer to a variable where the tristimulus Y will be stored	By reference
fZ	FLOAT *	pointer to a variable where the tristimulus Z will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.20 JETI_DWLPEEx

Returns the dominant wavelength (DWL) and color purity (PE) determined by the last measure. The calculation is based on a 2° observer, and the wavelength range for calculation is 380...780 nm.

The unit for the dominant wavelength is [nm].

The unit for the color purity is [%] (percent).

Prototype

DWORD JETI_DWLPEEx (DWORD_PTR dwDevice, FLOAT *fDWL, FLOAT *fPE)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fDWL	FLOAT *	Pointer to a variable where the dominant wavelength will be stored	By reference
fPE	FLOAT *	Pointer to a variable where the color purity will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.21 JETI_CCTEx

Returns the correlated color temperature (CCT) determined by the last measure. The wavelength range for calculation is 380...780 nm.

Prototype

DWORD JETI_CCTEx (DWORD_PTR dwDevice, float *fCCT)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fCCT	float *	Pointer to a variable where the correlated color temperature value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.22 JETI_DuvEx

Returns the Δuv for the correlated color temperature (CCT) determined by the last measure. The wavelength range for calculation is 380...780 nm.

Prototype

DWORD JETI_DuvEx (DWORD_PTR dwDevice, FLOAT *fDuv)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fDuv	FLOAT *	pointer to a variable where the Δuv value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.23 JETI_CRIEx

Returns the color rendering indices according to CIE 13.3-1995 publication and the color rendering index of the JIS color sample determined by the last measure.

The color temperature of the reference source is specified by fCCT. To use the same CCT as calculated, set fCCT to zero (0).

The function returns an array of 17 values containing the different CRI-values.

The first value (index 0) contains the chromaticity difference (DC) between the lamp to be tested and the reference illuminant. If DC is greater than 0.0054 the resulting color rendering indices may become less accurate.

The second value (index 1) contains the general color rendering index which is the arithmetical mean of eight special color rendering indices for the CIE-1974 test-color samples Nos. 1...8.

Value number three (index 2) to value number 17 (index 16) contains the special color rendering indices.

Prototype

DWORD JETI_CRIEx (DWORD_PTR dwDevice, float fCCT, float *fCRI)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fCCT	float	color temperature of the reference light source in [K]	By value
fCRI	float *	Pointer to an array where the CRI-values will be stored (the array must provide space for at least 17 values)	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.24 JETI_PeakFWHMEx

Returns the peak wavelength of a spectrum and the belonging width (FWHM–Full Width at Half of Maximum).

Prototype

DWORD JETI_PeakFWHMEx (DWORD_PTR dwDevice, float ffThreshold, float *fPeak, float* fFWHM)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fThreshold	float	threshold in %	By value
fPeak	float *	Pointer to a variable where the peak wavelength will be stored	By reference
fFWHM	float *	Pointer to a variable where the FWHM value will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.25 JETI_BlueMeasurementEx

Returns the results of calculations to assess the long-time effects of blue light radiation from displays on human eyes and the blue light ratios required to test the so called "low blue mode" of displays.

Prototype

DWORD JETI_BlueMeasurementEx (DWORD_PTR dwDevice, float* fLb, float* fKbv, float* fKc, float* fRbpfs, float* fRblbt, float* fRnbpbp)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fLb	float *	blue light hazard weighted radiance	By reference
fKbv	float *	blue light hazard efficacy	By reference
fKc	float *	circadian efficacy	By reference
fRbpfs	float *	ratio of a band with a bandwidth of +-20nm next to the blue peak to the entire power spectrum	By reference
fRblbt	float *	ratio of light in the range from 415nm - 455nm compared to 400nm – 500nm	By reference
fRnbpbp	float *	ratio of the highest peak outside the blue to the blue peak	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.26 JETI_RadioTintEx

This function gets the last used integration time.

Prototype

DWORD JETI_RadioTintEx (DWORD_PTR dwDevice, float *fTint)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
fTint	float *	Pointer to a variable where the integration time will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

3.27 JETI_SetMeasDistEx

This function sets the measuring distance which is used to calculate the values in intensity measuring mode.

Prototype

DWORD JETI_SetMeasDistEx (DWORD_PTR dwDevice, DWORD dwDistance)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
dwDistance	DWORD	the measuring distance in [mm]	By value

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS
0x...	see Appendix A for error codes

3.28 JETI_GetMeasDistEx

This function gets the measuring distance which is used to calculate the values in intensity measuring mode.

Prototype

DWORD JETI_GetMeasDistEx (DWORD_PTR dwDevice, DWORD *dwDistance)

Parameters

Input

Name	Type	Description	Call
dwDevice	DWORD_PTR	Handle to a device as returned by JETI_OpenRadioEx	By value
dwDistance	DWORD *	pointer to a variable where the measuring distance in [mm] will be stored	By reference

Return Value

Type	Description
DWORD	0x00 JETI_SUCCESS 0x... see Appendix A for error codes

4 Examples

To help starting development the SDK includes several examples.

4.2 C Examples

RadioSample

This sample demonstrates the basic usage of the jeti_radio DLL.

SyncSample

The SyncSample demonstrates the use of special functions to synchronize the measurements integration time with the frequency of pulsed light sources and pulsed monitor back-lights.

TriggerSample

This sample demonstrates the handle of measurements initiated by an external trigger event.

4.3 LabVIEW Examples

These samples demonstrate the basic usage of the DLLs within a LabVIEW program.

4.4 VisualBasic / VBA Examples

These sample demonstrate the usage of the jeti_radio DLL within a VBA macro inside an excel spreadsheet.

5 Appendix A

Error codes and their description:

<i>Error code</i>	<i>#define</i>	<i>Description</i>
0x00	JETI_SUCCESS	no error occurred
0x02	JETI_ERROR_OPEN_PORT	could not open COM-port
0x03	JETI_ERROR_PORT_SETTING	could not set COM-port settings
0x04	JETI_ERROR_BUFFER_SIZE	could not set buffer size of COM-port
0x05	JETI_ERROR_PURGE	could not purge buffers of COM-port
0x06	JETI_ERROR_TIMEOUT_SETTING	could not set COM-port timeout
0x07	JETI_ERROR_SEND	could not send to device
0x08	JETI_TIMEOUT	communication timeout error
0x0A	JETI_ERROR_RECEIVE	could not receive from device
0x0B	JETI_ERROR_NAK	command not supported or invalid argument
0x0C	JETI_ERROR_CONVERT	could not convert received data
0x0D	JETI_ERROR_PARAMETER	invalid argument
0x0E	JETI_BUSY	device busy
0x11	JETI_CHECKSUM_ERROR	invalid checksum of received data
0x12	JETI_INVALID_STEPWIDTH	invalid step width
0x13	JETI_INVALID_NUMBER	invalid device number
0x14	JETI_NOT_CONNECTED	device not connected
0x15	JETI_INVALID_HANDLE	invalid device handle
0x16	JETI_INVALID_CALIB	invalid calibration file number
0x17	JETI_CALIB_NOT_READ	calibration data not read
0x20	JETI_OVEREXPOSURE	measurement failed due to overexposure
0x22	JETI_MEASURE_FAIL	measurement failed due to other reasons
0x23	JETI_ADAPTION_FAIL	adaption failed
0x80	JETI_DLL_ERROR	internal DLL error
0xFF	JETI_FATAL_ERROR	fatal communication error
0x2710...0x2AFC		Windows sockets error codes

If a fatal communication error occurs (error code 0xFF) there are several ways to solve the problem.

1) Call JETI_HardReset (from jeti_core) to perform a device hardware reset. The effect of this function is the same as disconnecting then reconnecting the device from USB. This will work only if the device uses an FTDI USB-to-serial converter and was opened with direct access to the FTDI driver (opened with JETI_GetNumRadioEx and JETI_OpenRadioEx) instead of using the VCP (virtual com port) driver (JETI_OpenCOMDevice and/or JETI_SetComSearch).

Please note that all custom settings (e.g. integration time, function,...) will be set to default values and have to be repeated.

2) Closing the device with JETI_CloseRadioEx will also perform a hardware reset if a fatal communication error occurred on a device with FTDI USB-to-Serial converter. After closing the device it should be possible to reopen the device with JETI_GetNumRadioEx and JETI_OpenRadioEx.

3) If a JETI device with FTDI USB-to-Serial converter was opened using VCP driver (e.g. by using JETI_OpenCOMDevice) or by using other connections (like RS232, bluetooth,...) a fatal communication error can only be resolved by closing the device with JETI_CloseRadioEx and manually reset the device.

6 License Agreement

This End-User License Agreement ("EULA") is a legal agreement between you and JETI Technische Instrumente GmbH ("JETI") for the JETI SDK product which may include associated software components, media, printed materials, and "online" electronic documentation ("SOFTWARE PRODUCT").

You agree to be bound by the terms of this EULA by using the SOFTWARE PRODUCT. If you do not agree, do not use it.

1. COPYRIGHT

All rights, title, interests in and to copyrights in the SOFTWARE PRODUCT are owned exclusively (or licensed) by JETI. The SOFTWARE PRODUCT is protected by copyright laws and international copyright treaties, as well as other intellectual property laws and treaties. The SOFTWARE PRODUCT is licensed, not sold.

2. GRANT OF LICENSE

JETI grants Customer the following rights:

- a. Customer is granted the non-transferable, non-exclusive, and perpetual right to integrate the SOFTWARE PRODUCT in their software products, to use these products in their own company and to distribute these products under their own trade name.
- b. Customer is NOT required to pay any distribution, runtime, royalty, per-developer or per-end-user license fees.
- c. Customer may make copies of the SOFTWARE PRODUCT as may be necessary for backup and archival purposes.

3. RESTRICTIONS

- a. Customer may not disassemble, decompile, reverse engineer the SOFTWARE PRODUCT.
- b. Customer may not transfer, modify, sell, lease or sublicense the SOFTWARE PRODUCT.

4. TERMINATION

Without prejudice to any other rights, JETI may terminate this EULA if Customer fails to comply with the terms and conditions of this EULA. In such event, Customer must destroy all copies of the SOFTWARE PRODUCT in your possession.

5. NO WARRANTIES

JETI expressly disclaims any warranty for the SOFTWARE PRODUCT. JETI SDK is provided "As Is" without any express or implied warranty of any kind, including but not limited to any warranties of merchantability, noninfringement, or fitness of a particular purpose. JETI does not warrant or assume responsibility for the accuracy or completeness of any information, text, algorithms, graphics, links or other items contained within this SOFTWARE PRODUCT. JETI will not be liable for data loss, damages, loss of profits or any other kind of loss while using or misusing this library.

7 Service

In case of any questions or technical problems please contact:

JETI Technische Instrumente GmbH Göschwitzer Str. 48 D-07745 Jena Tel. +49 3641 23292 00 Fax +49 3641 23292 01 e-mail: sales@jeti.com Internet: www.jeti.com
--

Copyright (c) 2024 JETI Technische Instrumente GmbH. All rights reserved.

Software and operating instruction are delivered with respect to the License agreement and can be used only in accordance with this License agreement.

The hard and software as well as the operating instruction are subject to change without notice. JETI Technische Instrumente GmbH assumes no liability or responsibility for inaccuracies and errors in the operating instruction.

It is not allowed to copy this documentation or parts of it without previous written permission by JETI Technische Instrumente GmbH.